

AIRGAS TECHNOLOGY SDN BHD

+6012-712 5188

405660D

+6017-780 1118, +6017-880 5118

airgas.sales@gmail.com

www.airgas.my

No 19, Jalan EkoPernigaan 3, Kota Masai, 81750, Johor Bahru, Johor, Malaysia

Airgas



Gases



Gas Equipment



Cut & Weld Products



Personal Protection



Fire Fighting

MEDICAL AIR

MEDICAL GASES SPECIFICATION

IMPURITIES (V/V)	MEDICAL AIR
Min Purity (V/V)	≥ 99.9%
Hydrocarbon (V/V)	< 10 ppm
Odor	No Odor
Moisture (V/V)	≤ 100 ppm
Carbon Monoxide (V/V)	≤ 5 ppm
Carbon Dioxide (V/V)	≤ 5 ppm
Oxygen (V/V)	21% +/- 0.5%
Nitrous Oxide (V/V)	

AIR: THERE IS NO EUROPEAN PHARMACOPOEIA SPECIFICATION FOR MEDICAL COMPRESSED AIR. ALL COMPRESSED AIR FOR MEDICAL CYLINDER FILLING IS SUPPLIED TO THE FOLLOWING MODIFICATION

Refer to Requirements from the United States Pharmacopeia (USP)

Prepare on 1 Jan 2020